



Ref. Certif. No.
US/7379/ITS

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Digital Pressure/Temperature Gauges
Name and address of the applicant	Ralston Instruments LLC 15035 Cross Creek Parkway Newbury OH 44065 United States of America
Name and address of the manufacturer	Ralston Instruments LLC 15035 Cross Creek Parkway Newbury OH 44065 United States of America
Name and address of the factory	☐ Additional information on page 2 Ralston Instruments LLC 15035 Cross Creek Parkway Newbury OH 44065 United States of America
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	USB Mode - 5VDC, 50 mA Battery Mode: 3VDC, 50 mA Power Adapter: 100 ~ 240 VAC, 50/60Hz, 0.2A; Output: 5V @ 1A
Trademark (if any)	
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	LC10, LC20 or LC30; followed by four characters; followed by four characters; followed by two characters
Additional information (if necessary may also be reported on page 2)	☐ Additional information on page 2
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010/AMD1:2016
As shown in the Test Report Ref. No. which forms part of this Certificate	103405037COL-001

This CB Test Certificate is issued by the National Certification Body

Intertek Testing Services N.A., Inc.
545 E. Algonquin Road, Suite F, Arlington Heights IL 60005, United States of America

Date: 2018-09-07

Signature: John Quigley

page 1 of 1

원문: LC Gauge IEC / IECEE CB Test Certificate (61010-1) (assets.cms.ralstoninst.com) · 제조사 원문 자료입니다. 사양은 제조사 사정에 따라 변경될 수 있습니다.